

MX-RD-Pro LCD Digital Tube Rotator

Features

- LCD display of speed and time
- Circular rotating action, gives gentle but effective mixing
- Adjustable speed range from 10 to 70 rpm
- Adjustable angle from 0 to 90°
- Brushless motor ideal for extended use, and longer life
- Wide variety of accessories are available

Applications

Laboratory Tube Rotators are ideal for Biochemistry, Molecular Biology and Clinical applications. Provides a gentle but effective mixing of biological samples in 1.5mL to 50mL micro tubes, used in a variety of applications including immune precipitations, prevention of blood coagulation, latex diagnostics etc.





Specifications	MX-RD-Pro
Motor type	Brushless DC motor
Tilt angle	0-90°
Speed range	10-70rpm
Speed display	LCD
Timer display	LCD
Timer	1-1199min
Operation mode	Continuous / timer
Voltage,Frequency	100-240V, 50/60Hz
Power	40W
Dimension [D×W×H]	280×210×300mm
Permissible ambient temperature and humidity	5-40°C,80%RH
Protection class	IP21



Disk Accessories Type-1 Cat. No.18900160

Used for MX-RD-Pro for 1.5mLx60 centrifuge tubes holder



Disk Accessories Type-2 Cat. No.18900161

Used for MX-RD-Pro for 15mLx16 centrifuge tubes holder



Disk Accessories Type-3 Cat. No.18900162

Used for MX-RD-Pro for 50mLx8 centrifuge tubes holder







Rotisserie Accessory Type-3 Cat. No.18900144

Used for MX-RL-Pro.for 50mLx24 centrifuge tubes held horizontally



6

Disk support rods

Cat. No.18900140

4pcs in a group use with disk accessories



Rotisserie Accessory Type-4

Cat. No.18900145 Used for MX-RL-Pro.for 1.5mLx32

centrifuge tubes held vertically



Rotisserie Accessory Type-1 Cat. No.18900142

Used for MX-RL-Pro, for 1.5mLx48 centrifuge tubes held horizontally



Rotisserie Accessory Type-5 Cat. No.18900146

Used for MX-RL-Pro,for 15mLx16 centrifuge tubes held vertically



Rotisserie Accessory Type-2 Cat. No.18900143

Used for MX-RL-Pro, for 15mLx24 centrifuge tubes held horizontally



Rotisserie Accessory Type-6 Cat. No.18900147

Used for MX-RL-Pro, for 50mLx16 centrifuge tubes held vertically

